Quantifoil Instruments GmbH
Loebstedter Str. 101 . 07749 Jena . Germany
info@QInstruments.com +49 3641 876120
www.QInstruments.com

# EUROPEAN DECLARATION OF CONFORMITY

**Manufacturer name:** Quantifoil Instruments GmbH
**Address:** Loebstedter Str. 101 . 07749 Jena . Germany

We hereby explain that those corresponds to below designated products in its conception and design as well as in circulation meets the basic protection requirements relating to electromagnetic compatibility Directive and the Directive on the restriction of the use of certain hazardous substances.
In the case of a change of the product not coordinated with us this explanation loses its validity.

---

**Product type:** Heater shaker for laboratory
Heater for laboratory

| **Product name:** | | **with part no:** |
|---|---|---|
| | BioShake iQ | 1808-0506, 2016-0506 |
| | BioShake 3000-T | 1808-0516, 2016-0516 |
| | BioShake 3000-T elm | 1808-0517, 2016-0517 |
| | HeatPlate | 1807-0100, 2016-0100 |

---

**In accordance with relevant EC directives/standards:**

2014/30/EU The Electromagnetic Compatibility Directive
2011/65/EU Restriction of Hazardous Substances Directive
EN 61326-1:2013-07
DIN EN 61010-1:2011-07
DIN EN 61010-2-010:2015-05
DIN EN 61010-2-051:2004-07
EN 55011:2009 + A1:2010
EN 50581:2013-02 (based on)

---

**CE was at first applied:** 2011

---

**Released by:** Olaf Simmat
**Position:** Managing Director

**Date of issue:** January 1th, 2017
**Place of issue:** Jena, Germany

Olaf Simmat
.....
Signed by Olaf Simmat

Quantifoil Instruments GmbH
Loebstedter Str. 101 . 07749 Jena . Germany
info@QInstruments.com +49 3641 876120
www.QInstruments.com

# EUROPEAN DECLARATION OF CONFORMITY

**Manufacturer name:** Quantifoil Instruments GmbH
**Address:** Loebstedter Str. 101 . 07749 Jena . Germany

We hereby explain that those corresponds to below designated products in its conception and design as well as in circulation meets the basic protection requirements relating to electromagnetic compatibility Directive and the Directive on the restriction of the use of certain hazardous substances.
In the case of a change of the product not coordinated with us this explanation loses its validity.

---

**Product type:** Shaker for laboratory

| **Product name:** | | **with part no:** |
|---|---|---|
| | BioShake XP | 1808-0505, 2016-0505 |
| | BioShake 3000 | 1808-0016, 2016-0016 |
| | BioShake 3000 elm | 1808-0017, 2016-0017 |
| | BioShake 3000 elm DWP | 1808-0018, 2016-0018 |
| | BioShake 5000 elm | 1808-2017, 2016-2017 |

---

**In accordance with relevant EC directives/standards:**

2014/30/EU The Electromagnetic Compatibility Directive
2011/65/EU Restriction of Hazardous Substances Directive
EN 61326-1:2013-07
DIN EN 61010-1:2011-07
DIN EN 61010-2-051:2004-07
EN 55011:2009 + A1:2010
EN 50581:2013-02 (based on)

---

**CE was at first applied:** 2011

---

**Released by:** Olaf Simmat
**Position:** Managing Director

**Date of issue:** January 1th, 2017
**Place of issue:** Jena, Germany

.....
Signed by Olaf Simmat